



7.8 F-01



TEST REPORT

PARTY CODE : B/DLH/15104

SUBMITTED BY : BHUTANI INTERNATIONAL PVT. LTD.

295, Sant Nagar, East of Kailash, New Delhi-110065

SAMPLE NAME : SUPER UV STERILIZER LAMP

D. L. NO. : NS

MFG. BY : BHUTANI INTERNATIONAL PVT. LTD.

SUPP. BY : NS

REPORT NO. : FT0710000220

REF. NO. : NS

REF. DATE : 10/07/2020

DT.RECD : 10/07/2020

BATCH NO. : NS

BATCH SIZE : NS

SAMPLE QTY: 1 NO.

MFG. DT. : NS

EXP. DT. : NS



RESULTS OF ANALYSIS

Reference : Party Requirement

Description : One sample of "UV Sterilizer Lamp" was received by us on 10/07/2020.

S.No.	Parameter	Pathogens	Item	Results				
				Before UV (cfu/swab)	Sanitization Duration (Min)	After UV (cfu/swab)	Efficacy (%)	Method
1	Bacterial Count	Escherichia coli	Electronic Equipment and surface	560000	3	210	99.96	IS:5402:2012
				560000	5	120	99.98	IS:5402:2012
				560000	10	80	99.99	IS:5402:2012
				560000	15	40	99.99	IS:5402:2012

S.No.	Parameter	Pathogens	Item	Results				
				Before UV (cfu/swab)	Sanitization Duration (Min)	After UV (cfu/swab)	Efficacy (%)	Method
1	Fungus Count	Candida albicans	Electronic Equipment and surface	550000	3	250	99.95	IS:5402:2012
				550000	5	110	99.98	IS:5402:2012
				550000	10	80	99.99	IS:5402:2012
				550000	15	50	99.99	IS:5402:2012

Date of performance : 10/07/2020 to 15/07/2020

Note :- This product is found 99.99% effective against germs (Bacteria & Fungus)

DATE : 15/07/2020

PERSON IN CHARGE

REMARKS	Page: 1 of 1
1. Sample not drawn by us.	
2. Results listed refer only to the tested sample & applicable parameters. Endorsement of products is neither inferred nor implied.	
3. Samples will be destroyed after one month from the date of issue of test certificate unless otherwise specified.	
4. Liability of laboratory is limited to the invoiced amount only. Any dispute arising out of this report shall be subject to Delhi jurisdiction Only.	

5. This Report is not to be reproduced wholly or in part and cannot be used as an evidence in the Court of law and should not be used in any advertising media without special permission from the laboratory.

HARIDWAR LABORATORY : Plot No. 37, Sector - 7, I.I.E. SHDCUL, Haridwar - 249403

BUILDING MATERIAL LAB : Plot No. 107/11/2 & 107/20, Udhvoo Nagar Indl Area, Mundka, New Delhi-110041

Phone : +(91)-(01334)-235552

Phone : +(91)-93198-28884

Email : simahandwar@simalab.co.in

Email : bmlab@simalab.com

THIS CERTIFICATE IS NOT VALID WITHOUT A HOLOGRAM

SIMA